

**Burettes - Glass Key - Amber Graduation** 1

Same as above with Amber Graduation. CLASS-A. Accuracy as per DIN ISO 385 standard.

Class-A

Code	Capacity	Sub. Div.	Tolerance ±
CH0236I	10 ml	0.05 ml	0.03 ml
CH0236J	25 ml	0.10 ml	0.05 ml
CH0236K	50 ml	0.10 ml	0.05 ml
CH0236L	100 ml	0.20 ml	0.10 ml




**Burettes - PTFE Needle Valve**

With P.T.F.E. Needle Valve stopcock, blue/white graduation. Accuracy as per DIN ISO 385 standard.

Class-A

Blue Grad.	White Grad.	Capacity	Sub. Div.	Tolerance ±
CH0238A	CH0238A-WT	10 ml	0.05 ml	0.03 ml
CH0238B	CH0238B-WT	25 ml	0.10 ml	0.05 ml
CH0238C 2	CH0238C-WT	50 ml	0.10 ml	0.05 ml
CH0238D	CH0238D-WT	100 ml	0.20 ml	0.10 ml



Class-B

Blue Grad.	White Grad.	Capacity	Sub. Div.	Tolerance ±
CH0238E	CH0238E-WT	10 ml	0.05 ml	0.05 ml
CH0238F	CH0238F-WT	25 ml	0.10 ml	0.10 ml
CH0238G	CH0238G-WT	50 ml	0.10 ml	0.10 ml
CH0238H	CH0238H-WT	100 ml	0.20 ml	0.20 ml



High Accuracy



**Burettes - PTFE Key Stopcock**

With straight bore interchangeable PTFE Key stopcock. Accuracy as per DIN ISO 385 standard.

Class-A

Blue Grad.	White Grad.	Capacity	Sub. Div.	Tolerance ±
CH0240A	CH0240A-WT	10 ml	0.05 ml	0.03 ml
CH0240B	CH0240B-WT	25 ml	0.10 ml	0.05 ml
CH0240C	CH0240C-WT	50 ml	0.10 ml	0.05 ml
CH0240D	CH0240D-WT	100 ml	0.20 ml	0.10 ml

Class-B

Blue Grad.	White Grad.	Capacity	Sub. Div.	Tolerance ±
CH0240E	CH0240E-WT	10 ml	0.05 ml	0.05 ml
CH0240F	CH0240F-WT	25 ml	0.10 ml	0.10 ml
CH0240G 1	CH0240G-WT	50 ml	0.10 ml	0.10 ml
CH0240H	CH0240H-WT	100 ml	0.20 ml	0.20 ml